


CERTIFICATE OF ANALYSIS

General information

Name	18P-syndrome (46,XX,del(18)(p11.2)) Reference Standard
Cat. No.	CBPJ0008
Lot No.	CBFA23082406
Format	Genomic DNA
Genotype	18P-syndrome (46,XX,del(18)(p11.2))
Buffer	Tris-EDTA
Storage Conditions	2~8°C
Date of Manufacture	2023/08/24
Expiry	36 months from the date of manufacture

Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	40.00 ng/uL±10%	43.20 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.88 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>M: Marker 1: 18P-syndrome (46,XX,del(18)(p11.2)) Reference Standard</p>	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by: 彭亚楠 Date:2025/04/27

刘明 Date:2025/04/27