


CERTIFICATE OF ANALYSIS

General information

Name	18Q-syndrome (46,XX,del(18)(q22)) Reference Standard
Cat. No.	CBPJ0013
Lot No.	CBFA23122102
Format	Genomic DNA
Genotype	18Q-syndrome (46,XX,del(18)(q22))
Buffer	Tris-EDTA
Storage Conditions	2~8°C
Date of Manufacture	2023/12/21
Expiry	36 months from the date of manufacture

Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*2 vials	1.00 µg/vial*2 vials	√
Concentration	40.00 ng/uL±10%	42.40 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.76 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>M: Marker 1: 18Q-syndrome (46,XX,del(18)(q22)) Reference Standard</p>	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by: 彭亚楠 Date:2024/12/13

孙一柳 Date:2024/12/13