


CERTIFICATE OF ANALYSIS

General information

Name	HRD-P15-N (Normal)
Catalog No.	CBP90023-15
Lot No.	CGFA23040407
Format	Genomic DNA
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	2023/04/04
Expiry	36 months from the date of manufacture

Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	40.00 ng/uL±10%	39.40 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.87 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>M: Marker 1: HRD-P15-N</p>	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by: 陈毅 Date:2024/07/23

孙一扬 Date:2024/07/23